









# UV / VIS Spectrophotometer

Single Beam  
Double Beam  
NANO



# Product Overview

|                                  |  Page 4 |  Page 5 |  Page 6 |  Page 7 |  Page 8 |  Page 9 |  Page 10 |  Page 11 |
|----------------------------------|--|---|--|--|--|--|---|---|
| Model                            | EMC-11D-V  | EMC-11S-V   | EMC-11-UV  | EMC-11S-UV   | EMC-18S-UV   | EMC-61PC-UV  | EMC-61PCS-UV  | EMC-NANO-UV   |
| Single beam                      | ✓  | ✓   | ✓  | ✓  | ✓  |  |   | ✓   |
| Double beam                      |  |   |  |  |  | ✓  | ✓   |   |
| Spectral bandwidth               | 4 nm   | 4 nm  | 4 nm   | 4 nm   | 2 nm   | 1.8 nm   | 0.5/1/2/4/5 nm  | 4 nm  |
| Wavelength range                 | 325-1000 nm  | 325-1000 nm   | 200-1000 nm  | 200-1000 nm  | 190-1100 nm  | 190-1100 nm  | 190-1100 nm   | 190-1100 nm   |
| Calibration certificate          | ✓  | ✓   | ✓  | ✓  | ✓  | ✓  | ✓   | ✓   |
| Cell holder positions            | 4  | 4   | 4  | 4  | 4  | 1+1  | 1+1   | 1   |
| Audit Trail software (page 15)   | ○  | ○   | ○  | ○  | ○  | ○  | ○   | ○   |
| IQ/OQ/PQ                         |  | ○   | ○  | ○  | ○  | ○  | ○   | ○   |
| <b>PC Software (Windows®)</b>    |  |   |  |  |  |  |   |   |
| BASIC software (page 18)         | ✓  |   | ✓  |  |  |  |   |   |
| PROFESSIONAL software (page 18)  |  |   | ○  |  |  |  |   |   |
| EASY UV BASIC software (page 18) |  | ✓   |  | ✓  | ✓  |  |   |   |
| EASY UV software (page 18)       |  | ○   |  | ○  | ○  |  |   |   |
| ANALYST software (page 19)       |  |   |  |  |  | ✓  | ✓   | ✓   |
| <b>Accessories (page 12)</b>     |  |   |  |  |  |  |   |   |
| 8-position cell changer          |  |   |  |  | ○  | ○  | ○   |   |
| Peltier/Sipper plug in           |  |   |  |  | ○  | ○  | ○   |   |
| Calibration filter set           | ○  | ○   | ○  | ○  | ○  | ○  | ○   | ○   |

| PC Software (Windows®)    | BASIC | PROFESSIONAL | EASY UV BASIC | EASY UV | ANALYST |
|---------------------------|-------|--------------|---------------|---------|---------|
| Photometric measurement   | ✓     | ✓            | ✓             | ✓       | ✓       |
| Kinetic measurement       | ✓     | ✓            | ✓             | ✓       | ✓       |
| Quantitative measurement  | ✓     | ✓            | ✓             | ✓       | ✓       |
| DNA/Protein               |       | ✓            |               | ✓       | ✓       |
| Wavelength scan           |       | ✓            |               | ✓       | ✓       |
| Dongle (USB-Software key) |       | ✓            |               | ✓       | ✓       |

✓ = included  
○ = optionally

EMCLAB Spectrophotometers are tested with DAkkS (German body of accreditation) certified UV/VIS Reference Materials NIST traceable:

- Photometric Accuracy
- Wavelength Accuracy
- Spectral Resolution
- Stray Light

All EMCLAB Spectrophotometers are supplied with EMCLAB Works Calibration Certificate

EMCLAB Instruments GmbH  
Kulturstrasse 55  
47055 Duisburg | Germany  
Fon: +49 203 418 59 410  
Fax: +49 203 415 04 809  
E-mail: [info@emc-lab.de](mailto:info@emc-lab.de)  
Web: [www.emc-lab.de](http://www.emc-lab.de)



## Features:

- LCD screen (128\*64 dots)
- Manual setting wavelength
- Auto Zero and Blank
- Sample compartment for different cell holders
- Incl. PC software BASIC (Quantitative, Kinetics, Photometry Measurement)
- Incl. EMCLAB Works Calibration Certificate



| Model                     | EMC-11D-V                          |
|---------------------------|------------------------------------|
| Wavelength Range          | 325-1000 nm                        |
| Spectral Bandwidth        | 4 nm                               |
| Optical System            | Single Beam, grating 1200 lines/mm |
| Wavelength Accuracy       | ±2 nm                              |
| Wavelength Repeatability  | 1 nm                               |
| Photometric Accuracy      | ≤±0.5 % T or ±0.003A@1A            |
| Photometric Repeatability | ≤0.2 % T                           |
| Photometric Range         | 0-200 % T, -0.3 - 3A, 0-1999 Conc  |
| Photometric Mode          | T, A, C, F                         |
| Stray Light               | 0.3 % T                            |
| Stability                 | ±0.002A/h@500 nm                   |
| Noise                     | 0.003A@500 nm                      |
| Detector                  | Silicone Photodiode                |
| Display                   | LCD 128*64 dots                    |
| Central beam height       | 15 mm                              |
| Standard Cell Holder      | 4-position cell holder 10x10 mm    |
| Light Source              | Tungsten lamp                      |
| Output                    | USB port & Parallel port (printer) |
| Power Requirement         | AC 85V~265V 50/60 Hz               |
| Dimensions (LxWxH)        | 440 x 340 x 200 mm                 |
| Weight                    | 8 kg                               |

## Features:

- Color TFT screen
- Self-check system
- Fast choose wavelength
- Auto setting wavelength
- Auto Zero and Blank
- Sample compartment for different cell holders
- Incl. PC software EASY UV BASIC (Quantitative, Kinetics, Photometry Measurement)
- Incl. EMCLAB Works Calibration Certificate

## Optionally available:

- PC software EASY UV (Quantitative, Kinetics, Wavelength Scan, Multi-Wavelength, DNA/Protein, Energy Scan)
- Windows® Pad
- IQ/OQ/PQ documentation
- Audit Trail PC software (FDA 21 CFR part 11 compliant)



| Model                     | EMC-11S-V                            |
|---------------------------|--------------------------------------|
| Wavelength Range          | 325-1000 nm                          |
| Spectral Bandwidth        | 4 nm                                 |
| Optical System            | Single Beam, grating 1200 lines/mm   |
| Wavelength Accuracy       | ±2 nm                                |
| Wavelength Repeatability  | 1 nm                                 |
| Photometric Accuracy      | ≤±0.5 % T or ±0.004A@1A              |
| Photometric Range         | 0-200 % T, -0.3 - 3A 0-1999 conc.    |
| Photometric Repeatability | ≤0.2 % T                             |
| Photometric Mode          | T, A, C, F                           |
| Stray Light               | 0.2 % T                              |
| Stability                 | ±0.004A/h@500 nm                     |
| Baseline Flatness         | ±0.001A (200-1000 nm)                |
| Noise                     | 0.003A@500 nm                        |
| Detector                  | Silicone Photodiode                  |
| Display                   | Color TFT screen                     |
| Central beam height       | 15 mm                                |
| Standard Cell Holder      | 4-position cell holder 10x10 mm      |
| Light Source              | Tungsten lamp                        |
| Output                    | USB port & Parallel port for printer |
| Power Requirement         | AC 85V~265V 50/60 Hz                 |
| Dimensions (LxWxH)        | 440 x 350 x 200 mm                   |
| Weight                    | 8 kg                                 |

## Features:

- LCD screen (128\*64 dots)
- Self-check system
- Auto setting wavelength
- Auto Zero and Blank
- Sample compartment for different cell holders
- Save the results
- Up to 200 methods & 100 standard curves can be stored
- Incl. PC software BASIC (Quantitative, Kinetics, Photometry Measurement)
- Incl. EMCLAB Works Calibration Certificate



## Optionally available:

- PC software PROFESSIONAL (Quantitative, Kinetics, Wavelength Scan, Multi-Wavelength, DNA/Protein, Energy Scan)
- Windows® Pad
- IQ/OQ/PQ documentation
- Audit Trail PC software (FDA 21 CFR part 11 compliant)

| Model                     | EMC-11-UV                          |
|---------------------------|------------------------------------|
| Wavelength Range          | 200-1000 nm                        |
| Spectral Bandwidth        | 4 nm                               |
| Optical System            | Single Beam, grating 1200 lines/mm |
| Wavelength Accuracy       | ±2 nm                              |
| Wavelength Repeatability  | 1 nm                               |
| Photometric Accuracy      | ≤0.5 % T or ±0.003A@1A             |
| Photometric Repeatability | ≤0.2 % T                           |
| Photometric Range         | 0-200 % T, -0.3 - 3A, 0-9999 Conc. |
| Stray Light               | 0.2 % T                            |
| Stability                 | ±0.002A/h@500 nm                   |
| Noise                     | 0.003A@500 nm                      |
| Detector                  | Silicone Photodiode                |
| Display                   | LCD 128*64 dots                    |
| Central beam height       | 15 mm                              |
| Standard Cell Holder      | 4-position cell holder 10x10 mm    |
| Light Source              | Tungsten & Deuterium lamp          |
| Output                    | USB port & Parallel port (printer) |
| Power Requirement         | AC 110/230V 50/60 Hz               |
| Dimensions (LxWxH)        | 490 x 376 x 220 mm                 |
| Weight                    | 14 kg                              |

## Features:

- Color TFT screen
- Self-check system
- Fast choose wavelength
- Auto setting wavelength
- Auto Zero and Blank
- Sample compartment for different cell holders
- Incl. PC software EASY UV BASIC (Quantitative, Kinetics, Photometry Measurement)
- Incl. EMCLAB Works Calibration Certificate

## Optionally available:

- PC software EASY UV (Quantitative, Kinetics, Wavelength Scan, Multi-Wavelength, DNA/Protein, Energy Scan)
- Windows® Pad
- IQ/OQ/PQ documentation
- Audit Trail PC software (FDA 21 CFR part 11 compliant)



| Model                     | EMC-11S-UV                           |
|---------------------------|--------------------------------------|
| Wavelength Range          | 200-1000 nm                          |
| Spectral Bandwidth        | 4 nm                                 |
| Optical System            | Single Beam, grating 1200 lines/mm   |
| Wavelength Accuracy       | ±2 nm                                |
| Wavelength Repeatability  | 1 nm                                 |
| Photometric Accuracy      | ≤±0.5 % T or ±0.005A@1A              |
| Photometric Range         | 0-200 % T, -0.3 - 3A 0-19999 conc.   |
| Photometric Repeatability | ≤0.2 % T                             |
| Photometric Mode          | T, A, C, F                           |
| Stray Light               | 0.2 % T                              |
| Stability                 | ±0.004A/h@500 nm                     |
| Baseline Flatness         | ±0.001A (200-1000 nm)                |
| Noise                     | 0.003A@500 nm                        |
| Detector                  | Silicone Photodiode                  |
| Display                   | Color TFT screen                     |
| Central beam height       | 15 mm                                |
| Standard Cell Holder      | 4-position cell holder 10x10 mm      |
| Light Source              | Tungsten & Deuterium lamp            |
| Output                    | USB port & Parallel port for printer |
| Power Requirement         | AC 85V~265V 50/60 Hz                 |
| Dimensions (LxWxH)        | 440 x 350 x 200 mm                   |
| Weight                    | 9 kg                                 |

## Features:

- Color TFT screen
- Self-check system
- Auto setting wavelength
- Auto Zero and Blank
- Sample compartment for different cell holders
- Save the results
- Up to 200 methods & 100 standard curves can be stored
- Incl. PC software EASY UV BASIC (Quantitative, Kinetics, Photometry Measurement)
- Incl. EMCLAB Works Calibration Certificate



## Optionally available:

- PC software EASY UV (Quantitative, Kinetics, Wavelength Scan, Multi-Wavelength, DNA/Protein, Energy Scan)
- Windows® Pad
- IQ/OQ/PQ documentation
- Audit Trail PC software (FDA 21 CFR part 11 compliant)

| Model                     | EMC-18S-UV                           |
|---------------------------|--------------------------------------|
| Wavelength Range          | 190-1100 nm                          |
| Spectral Bandwidth        | 2 nm                                 |
| Optical System            | Single Beam, grating 1200 lines/mm   |
| Wavelength Accuracy       | ±0.5 nm                              |
| Wavelength Repeatability  | 0.3 nm                               |
| Photometric Accuracy      | ≤±0.5 % T or ±0.005A@1A              |
| Photometric Range         | 0-200 % T, -0.3 - 3A 0-9999 conc.    |
| Photometric Repeatability | ≤0.2 % T                             |
| Photometric Mode          | T, A, C, F                           |
| Stray Light               | 0.05 % T                             |
| Stability                 | ±0.001A/h@500 nm                     |
| Baseline Flatness         | ±0.001A (200-1000 nm)                |
| Noise                     | 0.003A@500 nm                        |
| Detector                  | Silicone Photodiode                  |
| Display                   | TFT color screen                     |
| Central beam height       | 15 mm                                |
| Standard Cell Holder      | 4-position cell holder 10x10 mm      |
| Light Source              | Tungsten & Deuterium lamp            |
| Output                    | USB port & Parallel port for printer |
| Power Requirement         | AC 110/230V 50/60 Hz                 |
| Dimensions (LxWxH)        | 490 x 380 x 220 mm                   |
| Weight                    | 11 kg                                |

## Features:

- Double Beam
- LCD screen (320\*240 dots)
- Self-check system
- Auto setting wavelength
- Auto Zero and Blank
- Sample compartment for different cell holders
- Incl. USB memory stick for methods and results
- Incl. PC software ANALYST  
(Quantitative, Kinetics, Photometry Measurement, Wavelength Scan, Multi-Wavelength, DNA/Protein, System Utilities)
- Incl. EMCLAB Works Calibration Certificate



## Optionally available:

- Windows® Pad
- IQ/OQ/PQ documentation
- Audit Trail PC software (FDA 21 CFR part 11 compliant)

| Model                     | EMC-61PC-UV  |
|---------------------------|--|
| Wavelength Range          | 190-1100 nm  |
| Spectral Bandwidth        | 1.8 nm   |
| Optical System            | Double Beam, grating 1200 lines/mm   |
| Wavelength Accuracy       | ±0.3 nm  |
| Wavelength Repeatability  | 0.2 nm   |
| Photometric Accuracy      | ≤±0.5 % T or ±0.005A@1A  |
| Photometric Range         | 0-200 % T, -0.3 - 3A   |
| Photometric Repeatability | ≤0.001A (0-0.5A), ≤0.002A (0-0.5A) ≤0.15 % T (0-100%)  |
| Scan Speed                | Hi, MED, LOW, MAX. 3000 nm/min   |
| Stray Light               | 0.05 % T   |
| Stability                 | ±0.001A/h@500 nm   |
| Baseline Flatness         | ±0.001A (200-1000 nm)  |
| Noise                     | 0.0003A@500 nm   |
| Detector                  | Silicone Photodiode  |
| Display                   | LCD 320*240 dots   |
| Central beam height       | 15 mm  |
| Standard Cell Holder      | 2 x Single cell holder 10x10 mm  |
| Light Source              | Tungsten & Deuterium lamp  |
| Output                    | USB port A for USB memory stick<br>USB port B for PC connectivity<br>Parallel port for printer |
| Power Requirement         | AC 110/230V 50/60 Hz   |
| Dimensions (LxWxH)        | 589 x 428 x 240 mm   |
| Weight                    | 22 kg  |



## Features:

- Double Beam
- Variable bandwidths 0.5/1/2/4/5 nm
- LCD screen (320\*240)
- Self-check system
- Auto setting wavelength
- Auto Zero and Blank
- Sample compartment for different cell holders
- Incl. USB memory stick for methods and results
- Incl. PC software ANALYST  
(Quantitative, Kinetics, Photometry Measurement, Wavelength Scan, Multi-Wavelength, DNA/Protein, System Utilities)
- Incl. EMCLAB Works Calibration Certificate



## Optionally available:

- Windows® Pad
- IQ/OQ/PQ documentation
- Audit Trail PC software (FDA 21 CFR part 11 compliant)

| Model                     | EMC-61PCS-UV   |
|---------------------------|--|
| Wavelength Range          | 190-1100 nm  |
| Spectral Bandwidth        | 0.5/1/2/4/5 nm   |
| Optical System            | Double Beam, grating 1200 lines/mm   |
| Wavelength Accuracy       | ±0.3 nm  |
| Wavelength Repeatability  | 0.2 nm   |
| Photometric Accuracy      | ≤±0.5 % T or ±0.005A@1A  |
| Photometric Range         | 0-200 % T, -0.3 - 3A   |
| Photometric Repeatability | ≤0.001A (0-0.5A), ≤0.002A (0-0.5A) ≤0.15 % T (0-100%)  |
| Scan Speed                | Hi, MED, LOW, MAX. 3000 nm/min   |
| Stray Light               | 0.05 % T@220, 340 nm   |
| Stability                 | ±0.001A/h@500 nm   |
| Baseline Flatness         | ±0.001A (200-1000 nm)  |
| Noise                     | 0.0003A@500 nm   |
| Detector                  | Silicone Photodiode  |
| Display                   | LCD 320*240  |
| Central beam height       | 15 mm  |
| Standard Cell Holder      | 2 x Single cell holder 10x10 mm  |
| Light Source              | Tungsten & Deuterium lamp  |
| Output                    | USB port A for USB memory stick<br>USB port B for PC connectivity<br>Parallel port for printer |
| Power Requirement         | AC 110/230V 50/60 Hz   |
| Dimensions (LxWxH)        | 589 x 428 x 240 mm   |
| Weight                    | 24 kg  |

## Features:


- LCD screen (320\*240 dots)
- Self-check system
- Auto setting wavelength
- Auto Zero and Blank
- Unique Flip cell holder 2 in 1 for NANO volume and standard cells
- 0.2 ~ 2.5  $\mu$ l sample volume for DNA, RNA, PCR, protein
- Incl. USB memory stick for methods and results
- Incl. PC software ANALYST  
(Quantitative, Kinetics, Photometry Measurement, Wavelength Scan, Multi-Wavelength, DNA/Protein, System Utilities)
- Incl. EMCLAB Works Calibration Certificate

## Optionally available:







- Windows® Pad
- IQ/OQ/PQ documentation
- Audit Trail PC software (FDA 21 CFR part 11 compliant)

## The unique Flip cell holder for NANO volume and standard cells:

- Simple to use
- NANO cell holder optical path length 0.5 mm
- Cell holder for standard cells optical path length 10 mm
- Horizontal light path - high precision - no loss of energy



| Model                     | EMC-NANO-UV  |
|---------------------------|--|
| Wavelength Range          | 190-1100 nm  |
| Spectral Bandwidth        | 4 nm   |
| Optical System            | Single Beam, grating 1200 lines/mm   |
| Wavelength Accuracy       | $\pm 0.5$ nm   |
| Wavelength Repeatability  | 0.3 nm   |
| Photometric Accuracy      | $\leq 0.5$ % T or $\pm 0.004A@1A$  |
| Photometric Repeatability | $\leq 0.001A$ (0-0.5A), $\leq 0.002A$ (0-0.5A) $\leq 0.5$ % T (0-100%)                         |
| Photometric Range         | 0-200 % T, -0.3 - 3A, 0-9999 Conc.   |
| Scan Speed                | Hi, MED, LOW – MAX: 3000nm/min   |
| Stray Light               | 0.05 % T@220, 340 nm   |
| Stability                 | $\pm 0.002A/h@500$ nm  |
| Baseline Flatness         | $\pm 0.002A$ (200-1000 nm)   |
| Noise                     | 0.0005A@500 nm   |
| Detector                  | Silicone Photodiode  |
| Display                   | LCD 320*240 dots   |
| Central beam height       | 15 mm  |
| Flip Cell Holder          | 0.2~2.5 $\mu$ l and cell 10 mm path length   |
| Light Source              | Tungsten & Deuterium lamp  |
| Output                    | USB port A for USB memory stick<br>USB port B for PC connectivity<br>Parallel port for printer |
| Power Requirement         | AC 110/230V 50/60 Hz   |
| Dimensions (LxWxH)        | 490 x 370 x 220 mm   |
| Weight                    | 14 kg  |

| Cell Holder   | Description   | Art. No.  |
|---|---|-----------|
|    | 4-position cell holder for cells 10x10 mm, without slider   | EMC-00020 |
|    | 4-position cell holder for cells up to 50 mm, without slider  | EMC-00021 |
|   | 4-position cell holder for cells up to 100 mm, without slider                                       | EMC-00019 |
|   | 1-position cell holder for cells 10x10 mm   | EMC-00023 |
|  | 1-position cell holder water-jacketed for cells 10x10 mm  | EMC-00024 |
|   | 4-position cell holder water-jacketed for cells 10x10 mm, without slider<br>(not for EMC-11 series) | EMC-00025 |
|  | 8-position auto cell changer for cells 10x10 mm<br>(not for EMC-11 series)                          | EMC-00029 |
|  | 1-position cell holder for cells up to 100 mm   | EMC-00045 |
|  | 1-position cell holder for solid samples up to 2-5 mm thickness                                     | EMC-00099 |
|   | 2-position cell holder for solid samples up to 2-5 mm thickness                                     | EMC-00101 |





| Peltier/Sipper System   | Description   | Art. No.  |
|---|---|-----------|
|  | Peltier Sipper System EMC-PSA 1, 10°C - 60°C,<br>incl. thermostat-controlled 1-position cell holder and<br>flow through cell 10x10 mm (not for EMC-11 series) | EMC-00125 |
|   | Peltier System EMC-PSA 2, 10°C - 60°C,<br>incl. thermostat-controlled 1-position cell holder<br>for cells 10x10 mm (not for EMC-11 series)                    | EMC-00148 |
|   | Sipper System EMC-PSA 3, incl. 1-position cell holder<br>and flow through cell 10x10 mm (not for EMC-11 series)   | EMC-00149 |

Fig. EMC-PSA 1

| Lamps   | Description   | Art. No.  |
|---|---|-----------|
|  | Halogen (Tungsten) lamp 6V10W (only for EMC-11D-V & EMC-11S series) | EMC-00011 |
|   | Halogen (Tungsten) lamp 12V/20 W (only for EMC-11-UV & EMC-18S-UV)  | EMC-00012 |
|   | Halogen (Tungsten) lamp 12V/20W (only for EMC-6 series)             | EMC-00013 |
|  | UV Deuterium lamp   | EMC-00039 |
|   | UV Deuterium lamp type 2 for EMC-11S-UV                             | EMC-00130 |

| Cells   | Description                           | Art. No. |
|---|---------------------------------------|----------|
|  | Glass / Quartz glass cells on request |          |

## UV/VIS Calibration filter set with EMCLAB Works Calibration Certificate for maintenance and service

The EMC-SET-100 for testing the wavelength accuracy and photometric accuracy enable the user to check their measurement results. The EMC-SET-100 is supplied with EMCLAB Works Calibration Certificate of Secondary Spectrometric Calibration Standards.

### UV/VIS reference materials for testing:

- Wavelength accuracy
- Photometric accuracy



## Glass filter set for testing wavelength accuracy (W) and photometric accuracy (A):

| Art. No.    | Filter                                     | Wavelength (nm)                  |
|-------------|--|----------------------------------|
| EMC-SET-100 | Holmium Oxide glass filter H1              | W*: 279; 361; 453; 536; 638 nm   |
|             | Neutral density glass filter N2, 0.25 Abs. | A*: 440; 465; 546.1; 590; 635 nm |
|             | Neutral density glass filter N3, 0.5 Abs.  | A*: 440; 465; 546.1; 590; 635 nm |
|             | Neutral density glass filter N4, 1.0 Abs.  | A*: 440; 465; 546.1; 590; 635 nm |
|             | Empty filter mount N0                      |                                  |

\*A= Wavelength for absorbance

\*W= Wavelength for wavelength accuracy

## EMCLAB Audit Trail Software (Windows®)

provides secure access of storage results and features compliant with 21 CFR Part 11.

System access requires a user name and password which are assigned by the workgroup manager/administrator. Individual user access levels determine the access to administrative tools which include instrument configuration, analysis applications, user setup, setup and security policies as well as system and application history logs.

For every file version the recorded information is saved (time logged, time created, time last written, deleted, user and manually recorded comments).

### EMCLAB Audit Trail is made for

- Monitoring of file modification
- Documentation of changes made by the user
- In specific directories
- Including subdirectories



AUDIT TRAIL

EMCLAB instruments Germany - AUDIT TRAIL

Setup

file

- Data
- Gerät1
- Gerät2
- Neuer Ordner
- EMC-61PCS-UV
  - Neuer Ordner
  - 2018
- 2017
- EMC-NANO-UV
- EMC-BR-12

| time logged                            | time created        | time last write     | length | deleted                             | comment         |
|--|---------------------|---------------------|--------|-------------------------------------|-----------------|
| file: UQC1707001_UV60_UV600_1.QUA      |                     |                     |        |                                     |                 |
| 07.06.2018 22:45:17                    | 07.06.2018 22:42:23 | 05.06.2018 15:03:04 | 1116   | <input type="checkbox"/>            |                 |
| file: UQC1707001_F1_Holmiumoxide_1.sca |                     |                     |        |                                     |                 |
| 07.06.2018 22:49:29                    | 07.06.2018 22:42:23 | 07.06.2018 22:49:24 | 84224  | <input checked="" type="checkbox"/> |                 |
| 07.06.2018 22:45:17                    | 07.06.2018 22:42:23 | 05.06.2018 15:01:24 | 84223  | <input type="checkbox"/>            |                 |
| file: UQC1707001_F1-F4.QUA             |                     |                     |        |                                     |                 |
| 07.06.2018 22:50:38                    | 07.06.2018 22:42:23 | 07.06.2018 22:50:38 | 1137   | <input type="checkbox"/>            | Außentemperatur |
| 07.06.2018 22:49:53                    | 07.06.2018 22:42:23 | 07.06.2018 22:49:53 | 1137   | <input type="checkbox"/>            | Zusatz A        |
| 07.06.2018 22:45:17                    | 07.06.2018 22:42:23 | 05.06.2018 15:01:35 | 1126   | <input type="checkbox"/>            | Ursprung        |

[Refresh](#)

Comment    Außentemperatur

K4=0.000000  
r=1.000000  
[Wavelength]  
Num=5  
WL1=445.000000  
WL2=465.000000  
WL3=546.099976  
WL4=590.000000  
WL5=635.000000  
WL6=850.000000  
WL7=800.000000  
WL8=750.000000  
WL9=700.000000  
WL10=650.000000  
WL11=600.000000  
WL12=580.000000  
WL13=560.000000  
WL14=540.000000  
WL15=520.000000  
WL16=500.000000  
WL17=400.000000  
WL18=480.000000  
WL19=460.000000

# AUDIT TRAIL

Disclaimer

## EMC-11D-V

### Mode A - Absorbance

Measured values are displayed in absorbance (Abs).

### Mode T - Transmission

Measured values are displayed in transmission (%T).

### Mode C - Concentration


A standard sample with known concentration is used to calculate the concentration of the measured sample.

### Mode F - Factor

The coefficients K and B of the standard curve equation  $C=K \cdot A+B$  are used to calculate the concentration of the measured sample.



Display



Main menu

## EMC-11-UV

### Basic


Measurements of absorbance, transmission or concentration.  
Display and save up to 200 values.

### Quantitative

A standard sample with known concentration is used to calculate the concentration of the measured sample.

### Utility

Switch lamps on/off, adjust test mode, dark current refresh, wavelength reset, version information



Display



Main menu

## EMC-11S-V/UV

**A Absorbance measurement** Measure the absorbance value of the sample.

**T Transmissivity measurement** Measure the transmission of the sample.

**C Concentration measurement against a standard sample**  
A standard curve through zero is defined by measuring the absorbance value of a standard sample and entering the corresponding concentration. The concentration of the unknown sample is calculated using the standard curve.

**F Concentration measurement using the equation coefficients**  
A standard curve is defined by entering the coefficients K and B of the standard curve equation  $C=K \cdot A+B$ , and the standard curve is used to measure the concentration of the unknown sample.

**E Energy measurement** For the verification of light source, monochromator and detector.

**System** General system information



Display



Main menu



## EMC-18S-UV



**Photometry** Measure the absorbance or transmission value of the sample.



**Quantitation** The concentration of the unknown sample is calculated using a standard curve.

- Standard sample: Measure or enter the absorbance value of a standard sample and enter the corresponding concentration value to define a standard curve.
- Equation coefficients: A standard curve is defined by entering the coefficients K and B of the standard curve equation  $C=K*A+B$ .




**Kinetics** Measure the photometric value change with time of the sample.



**System** General system information and settings (e.g. turn lamps on/off, set date and time).



Display



Main menu

## EMC-6-series and NANO

**Main menu:** The on-board software includes the following functions: photometry (basic mode), quantitation, wavelength scan, kinetics, DNA/Protein, multiwavelength, system utility.

**Quantitation:** Define a standard curve using up to 10 standards and choose from 4 curve fitting methods:

- Linear function
- Linear function through zero
- Square function
- Cubic function

**Wavelength scan:** Choose scan intervals (0.5 to 5 nm), scan speed, as well as the photometric mode to display the spectrum (wavelength/absorbance or wavelength/transmission).

Post processing options include rescaling of axes, curve tracking and determining peaks.


**Kinetics:** Absorbance vs time diagrams are displayed in real time. Choose time intervals (0.1 to 60 sec.), delay time as well as the photometric mode to display the curve (absorbance/time or transmission/time).

Post processing options include rescaling of axes, curve tracking and determining the part of the curve required for the rate calculation. The reaction rate is calculated with a linear regression algorithm.


**DNA/Protein:** Determine DNA/protein concentration at 260 nm/280 nm or 260 nm/230 nm with optional subtracted absorbance at 320 nm.

Optionally other wavelengths and factors may be entered.


**Multi-wavelength:** Measure multiple samples with up to 10 wavelengths.




Main menu




Quantitative Measurement



Wavelength scan



Kinetics



WL: 260.0nm

DNA/Protein Test

| No. | Items | Result | Unit  |
|-----|-------|--------|-------|
| 1   | A2    | 0.243  | Abs   |
|     | Aref  | 0.035  | Abs   |
|     | C-DNA | 4.524  | ug/ul |
|     | C-Pro | 110.8  | ug/ul |
|     | Ratio | 1.059  |       |

Search | Scroll

Coeff | Mode | Unit | Default

DNA/Protein



## PC software BASIC for the series EMC-11

### PC software EASY UV Basic for the series EMC-11S and EMC-18S

PC software based on Microsoft Windows. The spectrophotometer can be controlled by the PC software through the built-in USB communication port, which adds new functions to the UV/VIS series and facilitates operation. The software includes the following functions:


**Quantitative:** Use up to 20 standards to define a standard curve and choose from 3 methods for curve fitting:

1. Linear function
2. Linear function through zero
3. Cubic function

**Kinetics:** For reaction rate calculation or time course scanning. Absorbance vs time diagrams are displayed in real time.

**Photometry:** The test result can be displayed in absorbance or transmission.

Audit Trail PC software compliant with FDA 21 CFR Part 11 optionally available.



## PC software PROFESSIONAL for EMC-11-UV

### PC software EASY UV for the series EMC-11S and EMC-18S

PC software based on Microsoft Windows. The spectrophotometer can be controlled by the PC software through the built-in USB communication port, which adds new functions to the UV/VIS series and facilitates operation. The software includes the following functions:

**Quantitative:** Establish a standard curve: Enter or measure 20 standards or enter the standard curve equation using coefficients. Choose from 3 curve fitting methods:

1. Linear function
2. Linear function through zero
3. Cubic function

**Wavelength scan:** Choose scan intervals (0.1 to 5 nm), scan speed as well as the photometric mode to display the spectrum (wavelength/absorbance or wavelength/ transmission). Post processing options include i.a. rescaling of axes, curve smoothing, determining peaks and calculating derivatives.




**Kinetics:** For reaction rate calculation or time course scanning. Absorbance vs time diagrams are displayed in real time. Choose time intervals (0.5 to 3600 sec.), delay time as well as the photometric mode to display the curve (absorbance/ time or transmission/time). Post processing options include rescaling of axes, curve tracking and derivation.

**Multi-wavelength:** Enter up to 20 wavelengths to test multiple samples.

**DNA/Protein:** Determine DNA/protein concentration at 260 nm/280 nm or 260 nm/230 nm with optional subtracted absorbance at 320 nm. Optionally other wavelengths and factors may be entered.

**Energy scan:** For the verification of light source, monochromator and detector.

Audit Trail PC software compliant with FDA 21 CFR Part 11 optionally available.



## PC software ANALYST for EMC-6 series & EMC-NANO

The Analyst PC software enhances the functions and data processing and expands the storage capacity. The software includes the following functions:

**Quantitative:** Select only one wavelength in the 'method' tab in order to quantitatively calculate the measurement results using a standard curve. Enter or measure up to 20 standards to define a standard curve or enter the function of the standard curve using coefficients. Choose from 4 curve fitting methods:

1. Linear function
2. Linear function through zero
3. Square function
4. Cubic function

**Wavelength scan:** Choose scan intervals (0.1 to 5 nm), scan speed as well as the photometric mode to display the spectrum (wavelength/absorbance or wavelength/ transmission).

Post processing options include i.a. rescaling of axes, curve smoothing, determining peaks and calculating derivatives.




**Kinetics:** Absorbance vs time diagrams are displayed in real time. Choose time intervals (0.1 to 60 sec.), delay time as well as the photometric mode to display the curve (absorbance/ time or transmission/time).



Post processing options include rescaling of axes, curve tracking and determining the part of the curve required for the rate calculation. The reaction rate is calculated with a linear regression algorithm.

**DNA/Protein:** Determination of the DNA/protein concentration of the sample using common methods. Optionally other wavelengths and factors may be entered.

**Multi-wavelength:** Up to 20 different wavelengths can be entered to measure a number of samples.

Audit Trail PC software compliant with FDA 21 CFR Part 11 optionally available.





EMCLAB Instruments GmbH  
 Kulturstrasse 55  
 47055 Duisburg | Germany  
 Fon: +49 203 418 59 410  
 Fax: +49 203 415 04 809  
 E-mail: [info@emc-lab.de](mailto:info@emc-lab.de)  
 Web: [www.emc-lab.de](http://www.emc-lab.de)

Technical changes reserved